

Lot# MT080124

Marsil Laboratories LLC
3816 N Meridian Ave, Ste 113
Oklahoma City, Oklahoma 73112
405-917-5517
yomi@marsillab.com
www.marsillab.com



Marsil Lab ID: 4699
Date Sampled: 8/20/2024
Sampled By: N/A
Date Received: 8/20/2024
Received By: SBO

Contact: David R. High

Address: Millennial Technologies
6701 N Bryant
Oklahoma City, OK 73121

Date of Analysis: 8/24/2024

Temp Received: 29.4°C


Report of Analysis

Quantitative Results

Lab ID	Sample ID	Sample Description	Test Requested	Test Method	Dilution	Raw Count	Result (CFU/g)
4699 - 1	02	CARDIO CLEANSE	Aerobic Plate Count	USP <61>	10 ²	1	1.0 X 10 ²
4699 - 1	02	CARDIO CLEANSE	Yeast & Mold	USP <61>	10 ²	0	< 10 ²
4699 - 2	01	TURINGA	Aerobic Plate Count	USP <61>	10 ⁴	63	6.3 X 10 ⁵
4699 - 2	01	TURINGA	Yeast & Mold	USP <61>	10 ²	TNTC*	> 10 ²


*TNTC = Too Numerous To Count

Marsil Laboratories LLC
3816 N Meridian Ave, Ste 113
Oklahoma City, Oklahoma 73112
405-917-5517
yomi@marsillab.com
www.marsillab.com



Qualitative Results

Lab ID	Sample ID	Sample Description	Test Requested	Test Method	Result	Pass/Fail
4699 - 1	02	CARDIO CLEANSE	Salmonella	USP <62>	Absent	Pass
4699 - 1	02	CARDIO CLEANSE	<i>Esherichia coli</i>	USP <62>	Absent	Pass
4699 - 2	01	TURINGA	Salmonella	USP <62>	Absent	Pass
4699 - 2	01	TURINGA	<i>Esherichia coli</i>	USP <62>	Absent	Pass



Susan Beall-Overton
Laboratory Supervisor

8/24/2024
Date of Report

Results in the above report relate only to the samples tested which were received in satisfactory condition. This report shall not be reproduced, except in full, without the written approval of the laboratory.